

Supplementary Table III. Concordance of HCMV IgG CMIA positivity with HCMV specific IgG by ELISA

HCMV IgG	CMIA			Concordance
ELISA	Positive	Negative	Total	95.65%
Positive	154	3	157	
Negative	7	0	7	
Total	161	3	164	

ELISA, enzyme-linked immunosorbent assay; CMIA, chemiluminescent microparticle immunoassay