



**Supplementary Fig. 2.** User-directed actions recommended in response to FSCAs. This bar graph illustrates the categories of user actions recommended across the FSCA data analysed. Product replacement/returns, and software updates emerged as the most frequent actions advised, followed by performance monitoring and recalibration procedures. This figure complements the analysis of the practical impact of FSCAs on device usage in clinical settings.